

The dental restorations used for your case were constructed of FDA approved 510k materials and hand crafted in our North American Digital Processing Center. This is your assurance of quality and safety in materials and construction.

Item	Product Name	Manufacturer	Country of Origin	FDA 510k	Description
Alloy, High Noble Metal	d.Sign 96	Ivoclar	Liechtenstein	K011023	Au 73.8%, Pt 8.5%, Pd 5.4%, Ag 8.98%, In 1.9%, <1% Nb, Mn, Fe, Ta, Re, Ru, Li
Alloy, High Noble Metal	Evolution Lite	Ivoclar	Liechtenstein	K902376	Au 40.3%, Pd 39.3%, Ag 9.2%, In 9.3%, Ga 1.8%, Ru <1.0%
Porcelain	GC Initial MC, AL and LF	GC America	United States	K040567	Ceramic Powder made from SiO ₂ , Al ₂ O ₃ , K ₂ O, Na ₂ O, CaO, BaO, B ₂ O ₃ , <1% LiO ₂ , ZrO ₂ , other oxides (CeO ₂ , CeF ₃)